

ATL Accurate Testing Labs, LLC

7950 Meadowlark Way • Coeur d'Alene, ID 83815 • Phone (208) 762 8378 • Fax (208) 762 9082

Web site: www.accuratetesting.com e-mail: mueller@accuratetesting.com

Environmental Chemical Testing Laboratory

TCLP Samples

PARAMETER	VOL.	CONTAINER	PRESERVATIVE	HOLD TIME
METALS	16 oz.	Glass	None	180 days
MERCURY	16 oz.	Glass	None	28 days
BASE/NEUTRAL & ACID COMPOUNDS	16 oz.	Glass	None	14 days
CHLORINATED PESTICIDES & HERBICIDES	16 oz.	Glass	None	14 days
VOLATILES	8 oz.	Glass	None	14 days