

# ATL Accurate Testing Labs, LLC

7950 Meadowlark Way • Coeur d'Alene, ID 83815 • Phone (208) 762 8378 • Fax (208) 762 9082

Web site: [www accuratetesting.com](http://www accuratetesting.com) e-mail: [mueller@accuratetesting.com](mailto:mueller@accuratetesting.com)

## Environmental Chemical Testing Laboratory

### Organic Analysis

PARAMETER	VOL.	CONTAINER	PRESERVATIVE	HOLD TIME
BASE/NEUTRAL & ACID SEMIVOLATILE	2000 ml	Am. Glass	None	7 Days to Ext. 40 after Ext.
CHLORINATED HERBICIDES	1000 ml	Am. Glass	None	7 Days to Ext. 40 after Ext.
CHLORINATED PESTICIDES & PCB'S	1000 ml	Am. Glass	None	7 Days to Ext. 40 after Ext.
TPH: TOTAL RECOVERABLE PETROLEUM HYDROCARBONS (GC)*	1000 ml	Am. Glass	HCl	28 Days
* DRO/Diesel Range no Preservative				
VOLATILES	2X40 ml	Glass	HCl *	14 Days
(8260, BTEX, 624) * Do not use for 624				