

ATL Accurate Testing Labs, LLC

7950 Meadowlark Way • Coeur d'Alene, ID 83815 • Phone (208) 762 8378 • Fax (208) 762 9082

Web site: www.accuratetesting.com e-mail: mueller@accuratetesting.com

Environmental Chemical Testing Laboratory

Storm Water Samples

PARAMETER	VOL.	CONTAINER	PRESERVATIVE	HOLD TIME
COD, TKN, PO ₄ -P	1000 ml	Plastic	H ₂ SO ₄ to pH<2	28 Days
METALS (Cu, Pb, Zn)	250 ml	Plastic	HNO ₃ to pH<2	6 Months
NO ₃ -N ⁽¹⁾ , TSS ⁽²⁾ , pH ⁽³⁾	1000 ml	Plastic	None	48 Hours ⁽¹⁾ , 7 Days ⁽²⁾ , Immediately ⁽³⁾
FOG	1000 ml	Glass	HCl to pH<2	28 Days